

Quotation Number:	RFQ202104/013
Description:	SUPPLY AND DELIVERY OF 2X CHEMICAL OXYGEN DEMAND (COD) DIGESTOR BLOCKS
Bid Closing Date:	23/04/2021 @12PM
Compulsory Briefing Session:	N/A
Plant Address:	Laboratory
Contact Person	Kgomotso Tladinyane
Contact Number	011 929 7014/072 176 5387
Validity period	90 Days

Bids are hereby invited for the: Supply and delivery of 2x Chemical Oxygen Demand(COD) digester blocks.

Specifications:

Applications -Waste water test

-Water quality test

Temperature range -150 C/120 Min(COD program)

-Accuracy:2C

-Timer 30/60/120min (Via fixed programme

Number of vials -25 holes

Diameter of vials -16mm

Power -350W(AC110V/AC220V)

Safety -Anti-heat PTFE plate on the surface of heating block

-Auto shut off when block temperature is over 220°C

Block must be able to shut off automatically with audio alarm when every programme is completed.

Brochure with specifications must be submitted with the quotation.

Please e-mail your Quotation to rfq@erwat.co.za

